Identification of Microplastics Using Raman Spectroscopy

AN_R02; Angela Flack



Introduction

"Microplastics have invaded virtually every crevice on Earth" – a recent headline from a National Geographic article that highlights the global issue of the volume of microplastics in our environment.¹

Microplastics are defined as small solid particles of less than 5 mm composed of polymers. Additionally, they may also contain additives, such as dyes, and impurities from manufacturing. They are either described as primary microplastics which are plastics intentionally made small (e.g. microbeads), or secondary microplastics which are discarded plastic waste which has been broken down in marine environments and by the sun.

Microplastics are found everywhere in the environment including oceans, fish, salt, air, rivers, bottled water, and tap water. They accumulate in water and are eaten by animals such a fish and shellfish. These microplastics are then subsequently found in food and water for human consumption. There is a lot of debate and research around how dangerous microplastics are to the human body and so far, the results are inconclusive.

The World Health Organisation (WHO) last year called for more research into microplastics stating that currently, microplastics in drinking water don't appear to pose a risk to public health.² However, this is mainly due to the limited information available and there is no conclusive evidence that microplastics do not pose a risk to human health.

Raman spectroscopy is an extremely sensitive technique and gives rapid results, ideal for distinguishing polymers, contaminants, and dyes. Its fingerprint type analysis offers great potential as a tool in identifying microplastics in marine environments. Commonly employed techniques for studying microplastics are often limited to particle sizes above 300 μ m. However, only particles smaller than 150 μ m can be absorbed into the human gut and therefore this size range could be of vital importance.³ The high spatial resolution offered by confocal Raman spectroscopy means even the smallest of microplastics can be interrogated.

This application note will investigate 3 polymers, polyethylene (PE), polytetrafluoroethylene (PTFE), and nylon-6, using the Edinburgh Instruments RM5 Raman Microscope. PE is one of the most manufactured polymers, with tens of millions of tonnes made globally each year, much of this being used in packaging. PE is one of the top three most commonly found plastics in global aquatic systems. PTFE is a fluoropolymer used frequently for its non-stick properties, for example in frying pans. PTFE shares a very similar structure to PE

however their synthesis and uses differ significantly. There have been several cases recently where perfluorooctanoic acid (one of the synthesis materials for PTFE) has been found in drinking water in the USA and Australia, this acid is a carcinogen. Nylon-6 is a polyamide which frequently finds use in the automotive industry and in fishing materials. Discarded fishing equipment provides an instant route for nylon-6 to enter the aquatic ecosystem. These 3 polymers were chosen for their abundance in the Earth's oceans.

Materials and methods

An RM5 Raman Microscope equipped with a 785 nm laser and a X10 objective was used for the identification of microplastics. Three polymers of known particle size were purchased from Sigma Aldrich; nylon-6, PTFE, and PE. Their particle size being in the range of 35-50 μ m. The polymers were added to water and filtered through a Whatman[®] ME membrane filter.



Figure 1: RM5 Raman Microscope

Results and Discussion

Raman spectra of polymers and filter

As a reference, the Raman spectra of bulk nylon-6, PE and PTFE polymers were first measured to identify their characteristic peaks, Figure 2.

APPLICATION NOTE

Identification of Microplastics Using Raman Spectroscopy



AN_R02; Angela Flack



Figure 2: Raman spectra of nylon-6, PE, and PTFE

The spectrum of the filter was then acquired, the filter is a mixed cellulose ester membrane and therefore we see a nitrocellulose spectrum, Figure 3.





Raman spectra of unknown water sample

An unknown microplastic water sample was simulated by adding 1 mg of each plastic to 500 mL of water. This sample was filtered and left to dry. The filter was then investigated under the Raman microscope, in total 25 particles were selected to have their Raman spectra measured. The spectra contained both peaks attributed to the polymer and the filter. Figures 4-6 detail an example of three particles Raman spectra and their white light image (particle 8, particle 20, and particle 25), which were identified as PE, nylon-6, and



Figure 4: Spectra and white light image from particle 8



Figure 5: Spectra and white light image from particle 20



Figure 6: Spectra and white light image from particle 25

The KnowltAll Raman spectral library was then used for the identification of the plastic particles, KnowltAll ID Expert comes with all

APPLICATION NOTE

Identification of Microplastics Using Raman Spectroscopy

AN_R02; Angela Flack



Edinburgh Instruments Raman spectrometers.⁴ KnowltAll Raman spectral library is a database containing over 25000 Raman spectra. The KnowltAll technology applies a best fit comparison of the measured spectra with those in its library. It carries out a series of optimised corrections to provide the best possible match against its database.

After uploading the spectra of all 25 particles into KnowltAll the database search provides the most likely match, and gives a score out of 100 with 100 being a perfect match. For example, particle 8 was identified as PE with an 89.96% match. For this particle, PE was identified individually with no likely other contributors. However, some of the particles could not be uniquely identified, this was due to the spectral influence from the filter. For particle 6 KnowltAll identified the spectrum as a combination of 2 Raman spectra; nitrocellulose and PTFE. It weighted the contribution from PTFE at 0.56 and nitrocellulose at 0.44. Overall, the match for the particle spectrum being this combination was 92.83%. The 25 spectra were identified either individually or in a mixed result with nitrocellulose by KnowltAll, detailed in Table 1.

Particle	Polymer	Match	Particle	Polymer	Match
1	PE	92%	14	PTFE	93%
2	PTFE	94%	15	PE	91%
3	PTFE	94%	16	Nylon-6	80%
4	PTFE	94%	17	PTFE	97%
5	PTFE	91%	18	PE	94%
6	PTFE	93%	19	PTFE	81%
7	PE	93%	20	Nylon-6	77%
8	PE	92%	21	Nylon-6	81%
9	Nylon-6	80%	22	PE	95%
10	Nylon-6	82%	23	Nylon-6	81%
11	PE	94%	24	Nylon-6	84%
12	Nylon-6	92%	25	PTFE	94%
13	PE	90%			

Table 1: Results from KnowItAll for all 25 particles

accurate and rapid identification for the smallest of particles discovered. Raman spectroscopy can clearly distinguish different polymers at very small particle sizes easily. This application note demonstrates how the RM5 Raman Microscope combined with the KnowltAll database can be used to identify three types of microplastics commonly found in marine environments.

References

- National, G. Microplastics have invaded virtually every crevice on Earth A collection of new research provides more clues about where and how microplastics are spreading. 1–14 (2020).
- World Health Organization. Microplastics in drinking-water. (2019).
- 3. Presence of microplastics and nanoplastics in food, with particular focus on seafood. EFSA J. 14, (2016).
- KnowltAll ID Expert. https://sciencesolutions.wiley.com/knowitall-spectroscopysoftware/.

Conclusion

As we learn more about the spread of microplastics throughout the planet and their potential dangers there is an increasing need for

Identification of Microplastics Using Raman Spectroscopy

AN_R02; Angela Flack





For more information, please contact:

+44 (0) 1506 425 300 sales@edinst.com www.edinst.com